

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.0584		0.00100	1.30	1	06/23/2026 02:24	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:24	WG2777037	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.141		0.00100	1.30	1	06/23/2026 02:34	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:34	WG2777037	JDB

1
Cp

2
Tc

3
Ss

4
Cn

5
Sr

6
Qc

7
Gl

8
Al

9
Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.0700		0.00100	1.30	1	06/23/2026 02:37	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:37	WG2777037	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.120		0.00100	1.30	1	06/23/2026 02:40	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:40	WG2777037	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
--	--------	-------	-----------------

Analyte

Entry Point R

Sample Type A

Tier Number 1

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.245		0.00100	1.30	1	06/23/2026 02:44	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:44	WG2777037	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.0602		0.00100	1.30	1	06/23/2026 02:47	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:47	WG2777037	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.293		0.00100	1.30	1	06/23/2026 01:58	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:58	WG2777037	JDB

1
Cp2
Tc3
Ss4
Cn5
Sr6
Qc7
Gl8
Al9
Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.188		0.00100	1.30	1	06/23/2026 02:01	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:01	WG2777037	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.159		0.00100	1.30	1	06/23/2026 02:04	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:04	WG2777037	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.0459		0.00100	1.30	1	06/23/2026 02:08	WG2777037	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 02:08	WG2777037	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.118		0.00100	1.30	1	06/23/2026 01:12	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:12	WG2777041	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis date / time	Batch	Analyst
	mg/l		mg/l	mg/l				
Copper	0.166		0.00100	1.30	1	06/23/2026 01:22	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:22	WG2777041	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

013-412 HIGHLAND DR.

Collected date/time: 06/09/26 06:45

SAMPLE RESULTS - 13

L1985187

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.0349		0.00100	1.30	1	06/23/2026 01:25	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:25	WG2777041	JDB

1
Cp2
Tc3
Ss4
Cn5
Sr6
Qc7
Gl8
Al9
Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
--	--------	-------	-----------------

Analyte

Entry Point R

Sample Type A

Tier Number 1

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.126		0.00100	1.30	1	06/23/2026 01:28	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:28	WG2777041	JDB

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.0835		0.00100	1.30	1	06/23/2026 01:32	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:32	WG2777041	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.270		0.00100	1.30	1	06/23/2026 01:35	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:35	WG2777041	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al³ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.108		0.00100	1.30	1	06/23/2026 01:45	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:45	WG2777041	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.155		0.00100	1.30	1	06/23/2026 01:48	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:48	WG2777041	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

019-212 PAUL DR.

Collected date/time: 06/10/26 06:38

SAMPLE RESULTS - 19

L1985187

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.110		0.00100	1.30	1	06/23/2026 01:51	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:51	WG2777041	JDB

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.111		0.00100	1.30	1	06/23/2026 01:55	WG2777041	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 01:55	WG2777041	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis date / time	Batch	Analyst
	mg/l		mg/l	mg/l				
Copper	0.417		0.00100	1.30	1	06/23/2026 03:03	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:03	WG2777045	JDB

1
Cp2
Tc3
Ss4
Cn5
Sr6
Qc7
Gl8
Al9
Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.176		0.00100	1.30	1	06/23/2026 03:13	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:13	WG2777045	JDB

1
Cp

2
Tc

3
Ss

4
Cn

5
Sr

6
Qc

7
Gl

8
Al

9
Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.146		0.00100	1.30	1	06/23/2026 03:16	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:16	WG2777045	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis date / time	Batch	Analyst
	mg/l		mg/l	mg/l				
Copper	0.0638		0.00100	1.30	1	06/23/2026 03:20	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:20	WG2777045	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
--	--------	-------	-----------------

Analyte

Entry Point R

Sample Type A

Tier Number 1

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.0316		0.00100	1.30	1	06/23/2026 03:23	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:23	WG2777045	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis date / time	Batch	Analyst
	mg/l		mg/l	mg/l				
Copper	0.110		0.00100	1.30	1	06/23/2026 03:26	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:26	WG2777045	JDB

¹ Cp² Tc³ Ss⁴ Cn⁵ Sr⁶ Qc⁷ Gl⁸ Al⁹ Sc

027-2201 EWING DR

Collected date/time: 06/10/26 06:20

SAMPLE RESULTS - 27

L1985187

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
--	--------	-------	-----------------

Analyte

Entry Point R

Sample Type A

Tier Number 1

Metals (ICPMS) by Method 200.8

Analyte	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
	mg/l		mg/l	mg/l		date / time		
Copper	0.349		0.00100	1.30	1	06/23/2026 03:36	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:36	WG2777045	JDB

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.0403		0.00100	1.30	1	06/23/2026 03:39	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:39	WG2777045	JDB

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.0486		0.00100	1.30	1	06/23/2026 03:42	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:42	WG2777045	JDB

1Cp

2Tc

3Ss

4Cn

5Sr

6Qc

7Gl

8Al

9Sc

Additional Information - Results for field analyses are not accredited to ISO 17025

	Result	Units	Reference Limit
Analyte			
Entry Point	R		
Sample Type	A		
Tier Number	1		

Metals (ICPMS) by Method 200.8

	Result	Qualifier	Det. Limit	Reference Limit	Dilution	Analysis	Batch	Analyst
Analyte	mg/l		mg/l	mg/l		date / time		
Copper	0.176		0.00100	1.30	1	06/23/2026 03:46	WG2777045	JDB
Lead	ND		0.00200	0.0150	1	06/23/2026 03:46	WG2777045	JDB

¹ Cp

² Tc

³ Ss

⁴ Cn

⁵ Sr

⁶ Qc

⁷ Gl

⁸ Al

⁹ Sc

WG2777037

Metals (ICPMS) by Method 200.8

QUALITY CONTROL SUMMARY

[L1985187-01,02,03,04,05,06,07,08,09,10](#)

Method Blank (MB)

(MB) R4390921-1 06/23/26 02:17

Analyte	MB Result mg/l	MB Qualifier	MB MDL mg/l	MB RDL mg/l
Copper	U		0.000270	0.00100
Lead	U		0.000212	0.00200

Laboratory Control Sample (LCS)

(LCS) R4390921-2 06/23/26 02:21

Analyte	Spike Amount mg/l	LCS Result mg/l	LCS Rec. %	Rec. Limits %	LCS Qualifier
Copper	0.0500	0.0488	97.7	85.0-115	
Lead	0.0500	0.0478	95.6	85.0-115	

L1985187-01 Original Sample (OS) • Matrix Spike (MS) • Matrix Spike Duplicate (MSD)

(OS) L1985187-01 06/23/26 02:24 • (MS) R4390921-3 06/23/26 02:27 • (MSD) R4390921-4 06/23/26 02:31

Analyte	Spike Amount mg/l	Original Result mg/l	MS Result mg/l	MSD Result mg/l	MS Rec. %	MSD Rec. %	Dilution	Rec. Limits %	MS Qualifier	MSD Qualifier	RPD %	RPD Limits %
Copper	0.0500	0.0584	0.104	0.105	91.9	93.4	1	70.0-130			0.707	20
Lead	0.0500	ND	0.0472	0.0470	93.3	92.9	1	70.0-130			0.344	20

1 Cp

2 Tc

3 Ss

4 Cn

5 Sr

6 Qc

7 Gl

8 Al

9 Sc